Notice of Allowability	Application No.	Applicant(s) CHARLES, STEVEN T.	
	10/672,188		
	Examiner	Art Unit	
	Laura A. Bouchelle	3763	
The MAILING DATE of this communication at All claims being allowable, PROSECUTION ON THE MERITS herewith (or previously mailed), a Notice of Allowance (PTOL NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATEN of the Office or upon petition by the applicant. See 37 CFR 1	S IS (OR REMAINS) CLOSED in th -85) or other appropriate communic T RIGHTS. This application is sub	is application. If not included cation will be mailed in due cours	se. THIS
1. This communication is responsive to <u>Amendment filed</u>	<u>6/12/07</u> .	÷.	
2. X The allowed claim(s) is/are 15-26.			
3.	have been received. have been received in Application Now documents have been received in TE" of this communication to file a son DNMENT of this application. The submitted of the attached EXAMITY of the submitted of	No I this national stage application for this national stage application for the reply complying with the requirer INER'S AMENDMENT or NOTICE claration is deficient. PTO-948) attached the Office action of drawings in the front (not the back 1.121(d). IAL must be submitted. Note:	ments CE OF
Attachment(s) 1. ☐ Notice of References Cited (PTO-892)	5. □ Notice of Infor	mal Patent Application	
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-9	48) 6. 🗌 Interview Sum	mary (PTO-413),	
3. Information Disclosure Statements (PTO/SB/08),		il Date nendment/Comment	
Paper No./Mail Date 4.		atement of Reasons for Allowand	ce
•	9. 🗌 Other		

Application/Control Number: 10/672,188 Page 2

Art Unit: 3763

Allowable Subject Matter

1. The following is an examiner's statement of reasons for allowance: The subject matter of the independent claims could not be found and was not suggested by the prior art of record. The subject matter not found is a method of draining fluid from the eye comprising the steps of providing a vitreoretinal instrument having a cannula with a curved distal portion, a first and second port, and a valve, and disposing the cannula in the eye, closing the second port, aspirating fluid through the first port, opening the second port, and aspirating fluid through both ports in combination with the features of the claim.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Laura A. Bouchelle whose telephone number is 571-272-2125. The examiner can normally be reached on Monday-Friday 8-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 517-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/672,188 Page 3

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Laura A Bouchelle

Examiner

Art Unit 3763

MICHOLAS D. LUCCKESI

SUPERVISORY PATELLE LE VALVINZA TECHNOLOGY CENTER 2700